

Name of Accreditation Program	JCSS Accreditation Program
Accreditation Identification	JCSS 0324 Calibration
Name of Conformity Assessment Body	TAMAGAWA TECHNO CREATION Co., Ltd.
Name of Legal Entity	TAMAGAWA TECHNO CREATION Co., Ltd. JCN 8100001023354
Inquiry Point	Calibration Operation Department TEL: +81-265-21-7661      FAX: +81-265-21- 7662

\*JCN: Japan Corporate Number



23・12・08-NITE-007  
2 0 2 4 - 0 5 - 3 1

# Certificate of Accreditation

International Accreditation Japan (IAJapan) hereby accredits the following conformity assessment body as a calibration laboratory of Japan Calibration Service System.

Accreditation Identification: JCSS 0324 Calibration

Name of Conformity Assessment Body: TAMAGAWA TECHNO CREATION Co.,Ltd.

Name of Legal Entity: Same as above

Location of Conformity Assessment Body: 1879 Ohyasumi, Iida-shi, Nagano 395-8515, JAPAN

Scope of Accreditation: Angle (as the following page)

Accreditation Requirement: ISO/IEC 17025:2017\*

\* The relevant accreditation requirements described in the Accreditation Scheme Document for JCSS are also applied.

Effective Date of Accreditation: 2024-06-01

Expiry Date of Accreditation: 2028-05-31

Date of Initial Accreditation: 2016-10-20

JCSS

Hideaki Tanaka

TANAKA Hideaki

Chief Executive, International Accreditation Japan (IAJapan)

National Institute of Technology and Evaluation

- International Accreditation Japan (IAJapan) is a laboratory accreditation body which has signed MRAs of ILAC (International Laboratory Accreditation Cooperation) and APAC (Asia Pacific Accreditation Cooperation).
- MRA requirements are, in addition to relevant international standards and guides, requirements for participation in proficiency testing programs, surveillance and reassessment, and the policy for the traceability of measurement for MRA purpose.
- This laboratory fulfills ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This accreditation means this laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).
- The latest accreditation information is publicly available on IAJapan Website as an accreditation certificate.

General Field of Calibration: AngleDate of Initial Accreditation of the Field: 2016-10-20Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facilityCalibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated	Range		Expanded Uncertainty (Level of Confidence Approximately 95 %)
Angle Measuring Instrument	Rotary Encoder	0 ° to 360 ° up to 36,000 points	0.062"

#All Calibration Procedures are in-house procedures developed by this laboratory.